

Medical Device White Paper Series

Software as a medical device

A comparison of the EU's approach
with the US's approach

Authors - Pat Baird, Head of Global Software Standards,
Philips, and **Koen Cobbaert**, Senior Manager Quality,
Standards and Regulations, Philips



Disclaimer – This white paper is issued for information only. It does not constitute an official or agreed position of BSI Standards Ltd. The views expressed are entirely those of the authors. All rights reserved. Copyright subsists in all BSI publications including, but not limited to, this white paper. Except as permitted under the Copyright, Designs and Patents Act 1988, no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. While every care has been taken in developing and compiling this publication, BSI accepts no liability for any loss or damage caused, arising directly or indirectly in connection with reliance on its contents except to the extent that such liability may not be excluded in law.

Introduction

The International Medical Device Regulators Forum (IMDRF) aims to accelerate international medical device regulatory convergence. It comprises representatives from the medical device regulatory authorities of Australia, Brazil, Canada, China, the EU, Japan, the Russian Federation, Singapore, South Korea and the US. Through the IMDRF, regulators reached consensus on what software is considered a medical device. Regulators call it 'software as a medical device' (SaMD).¹ As SaMD might be regulated in one country but not in another, this is an important consideration for manufacturers' go-to-market strategies and for the availability of SaMD across the world. This paper provides a comparison of how SaMD is regulated in the US and in the EU.

History

In 2007, the Swedish Medical Products Agency, Läkemedelsverket, analysed the top 10 incidents in health care. It noticed that software was often the root cause² – for example medical data being lost from a temporary database, medication being assigned to the wrong patient, the wrong dose being calculated, failure to caution that a patient is allergic to an active substance, etc. In several cases, software had caused the death of a patient. Läkemedelsverket investigated and found that a lot of the software used in health care qualified as a medical device but had not undergone a conformity assessment and did not carry CE marking. Even though EU regulators had long considered that software-only products could be subject to medical device legislation, it had not been very clear to people that the term 'device' could also be applicable to something intangible like software.




Figure 1 Date as of which countries recognized that software-only products could be medical devices. (Cobbaert, 2020)

Sweden took the issue forward when it held the presidency of the Council of the European Union. In 2009, it convinced other regulators to add the word 'software' to the EU definition of 'medical device'. In 2011, the Global Harmonization Task Force (GHTF), the predecessor of the IMDRF, also changed its definition. This caused a ripple effect across the world (see Figure 1). Following the direction of the World Health Organization (WHO), other countries started putting in place medical device legislation and leveraged the new GHTF definition for 'medical device', in effect regulating SaMD. Later, in 2013, IMDRF published its guidance to clarify what software it considers to be a medical device.

Software as a medical device

The IMDRF defines 'SaMD' as 'software intended to be used for one or more medical purposes that perform these purposes *without being part of* a hardware medical device'. The IMDRF provides clarification through notes, further supplemented by U.S. Food & Drug Administration (FDA) guidance.³ An important clarification is that the italicized term does not refer to the physical location from where the software is running but to the regulatory status of the software. Software can run on general-purpose IT equipment in 'the cloud' but also on the computing platform of a hardware medical device and still be SaMD. When the hardware medical device needs the software to achieve its intended medical purpose – for example because it drives the hardware or fulfils a purpose claimed for the hardware device – then the software is not SaMD but part of the medical device in the regulatory meaning of the term. For example, consider software for automatic nerve detection intended to run on the computing platform of an ultrasound device. A manufacturer can place such software on the market as SaMD or as part of the ultrasound device, depending on whether the manufacturer wants to assign the nerve detection claim to the ultrasound device or just to the software. Software that does not fulfil a medical purpose on its own, on the other hand, is not SaMD. For example, software intended to solely drive an ultrasound transducer can be placed on the market as an integral part of the ultrasound device or as an accessory of the ultrasound device.

Placement on the market

Software can be qualified and placed on the market as:

- a medical device or in vitro diagnostic (IVD) medical device (the focus of this paper)
- an accessory for a medical device or for an IVD medical device (accessories by definition do not fulfill a medical purpose on their own)
- a part or a component of a medical device, IVD medical device or Annex XVI device (Annex XVI devices have no medical purpose but are in scope of the EU Medical Device Regulation (MDR⁴)).

If the software is none of the above, it is not subject to the EU MDR, the EU In Vitro Diagnostic Regulation (EU IVDR)⁵ or FDA regulations, unless it is placed on the EU market as part of a system – that is a combination of products, either packaged together or not, that are intended to be interconnected or combined or to achieve a specific medical purpose, in which case it is subject to the EU MDR. Connectivity alone is not sufficient for it to be considered a system,⁶ because there is a third condition: the system must be placed on the market as one unit – for example it is sold under a single sales catalogue number.

EU MDR and EU IVDR Article 6 imply that software not placed on the European market might still have to comply with the EU MDR if offered, directly or through intermediaries, to a person established in the EU. Think of software offered as a download or as a service through web portals and application interfaces. If such software operates on servers based outside the EU, then such software might nevertheless be subject to the EU MDR or EU IVDR if it is accessible through, for example, website subscription to a person residing in the EU. Two years after publication of the EU MDR, the FDA clarified that, in the US, software as a service might be regulated as a medical device.⁷

SaMD that is not regulated

For the purpose of this paper, 'software' is defined as 'a set of instructions that processes input data and creates output data'. Software that does not meet this definition is 'inactive' – for example digital libraries, medical models and orthopaedic templates. Even if such software has a medical purpose, it is generally not subject to the regulation unless it is placed on the market as a part of or a component of a medical device, IVD medical device or Annex XVI device; as an accessory for a medical device or IVD medical device; or as part of a system. Such software is not in scope of this paper.

The FDA uses the fault lines provided by the IMDRF SaMD risk framework⁸ to clarify what software is not in scope of its regulation, whereas the EU uses functional exemptions instead.


In the EU, SaMD with functionality that is limited to storage, communication, lossless compression or simple searching is not regulated.⁹

'Communication' refers to the transfer of data, data parsing (i.e. converting a string of data from one syntax to another, e.g. from a proprietary syntax to HL7), converting units (e.g. pound to kilogramme), and altering representation of information for embellishment or compatibility purposes.

'Lossless compression' refers to a compression procedure that allows the exact reconstruction of the original data.

'Simple searching' refers to the retrieval of records by matching record metadata against record search criteria or retrieval of information. Pattern detection and detecting whether a parameter is within bounds do not constitute simple searching.

In the US, SaMD to inform clinical management is not regulated if it is intended for a health care professional to independently review and understand the basis of the software recommendation on condition that the software does not perform signal or image acquisition, processing or analysis. Consequently, the EU regulation has a larger scope than that of the US, including significantly more clinical decisions support software intended to inform clinical management, such as drug–drug interaction and allergy checkers¹⁰ (see Figure 2).



SEROTONIN SYNDROME MIGHT CAUSE THE FOLLOWING SYMPTOMS:

Confusion, hallucination, seizure, extreme changes in blood pressure, increased heart rate, fever, excessive sweating, shivering or shaking, blurred vision, muscle spasm or stiffness, tremor, incoordination, stomach cramp, nausea, vomiting, diarrhoea and, in extreme cases, coma and death

Figure 2 Example of the output of a drug–drug interaction checker. Drug–drug interaction and allergy checkers are used as part of drug prescription software. These checkers use information from pharmacopoeia and the electronic patient file to issue warnings related to possible adverse reactions when prescribing a drug that is contraindicated for a drug already taken by the patient or to which the patient is allergic. Such software is regulated in the EU⁹ but not in the US.¹¹ (Cobbaert, 2020)

Finally, the EU and the US also do not regulate SaMD intended for conducting clinical investigations or population and epidemiological studies, because such software is not intended for providing diagnosis or treatment information for an individual, even though such software might use data from individuals – for example software to map the spread of COVID-19 in a country.

Regulatory requirements: FDA

The FDA has published multiple guidance documents regarding the regulation of software, including SaMD. Some types of software are regulated as medical devices, whereas other types of software are not regulated, and a third type of software is subject to ‘enforcement discretion’ – technically, the product is regulated but the FDA will not actively pursue enforcement unless there is a reason to.

Types of medical software

The FDA has defined stand-alone medical software in the Preamble Section AA Special Requirements for Stand-Alone Software—Final § 801.50 of the Unique Device Identification (UDI) System Regulations as ‘medical software that is itself a medical device and is not a component, part, or accessory of a medical device’.¹²

The FDA has also published additional guidance documents about specific types of software that are used in health care that might be regulated. For example, ‘Medical Device Data Systems’ (MDDS) are defined as:

... hardware or software products intended to transfer, store, convert formats, and display medical device data. A MDDS does not modify the data or modify the display of the data, and it does not by itself control the functions or parameters of any other medical device. MDDS may or may not be intended for active patient monitoring.¹³



These applications might or might not be regulated, depending on other functions of the application. In general, if the software is solely intended to transfer, store, convert and display, then it is not subject to regulation. Examples include software that stores patient data for review at a later time, software that converts data into a format that can be printed and software that displays a previously stored electrocardiogram for a patient.

Note that the software does not modify the data and the software does not control the functions or parameters of a device. If the software has those features, it is likely to be a device. For example, software intended to generate alarms or prioritize patient-related information on a display would likely be considered a medical device. Similarly, software that detects and highlight abnormalities (computer-assisted detection (CADe)) or software that assesses associated disease severity (computer-assisted diagnosis (CADx)) is considered a device by the FDA and is subject to regulatory focus.

Similarly, 'clinical decision support' (CDS) software is defined as follows:

Clinical decision support (CDS) provides clinicians, staff, patients or other individuals with knowledge and person-specific information, intelligently filtered or presented at appropriate times, to enhance health and health care. CDS encompasses a variety of tools to enhance decision-making in the clinical workflow. These tools include computerized alerts and reminders to care providers and patients; clinical guidelines; condition-specific order sets; focused patient data reports and summaries; documentation templates; diagnostic support, and contextually relevant reference information, among other tools.¹⁴

Because of the variety of CDS applications, as well as an evolving regulatory landscape, some CDS software might be regulated by the FDA, other CDS software might be regulated by the FDA but under 'enforcement discretion', and some CDS software might not be regulated as medical devices. Although the FDA released draft guidance³ on this topic in September 2019, final guidance has not been published.

Factors that affect whether or not the CDS software is regulated include:

- Is the intended user a health care provider (HCP)?
- Can the user independently review the basis of the CDS software information?
- What is the state of the health care situation or condition (is the patient state non-serious, serious or critical)?

Table 1 from the draft CDS guidance shows how these factors affect the device regulation:

		Intended user is HCP	Intended user is Patient or Caregiver
IMDRF Risk Categorization	Can the User Independently Review the Basis?*	FDA Regulation	FDA Regulation
Inform x Critical	Yes	Not a Device	Oversight Focus
	No	Oversight Focus	
Inform x Serious	Yes	Not a Device	
	No	Oversight Focus	
Inform x Non-Serious	Yes	Not a Device	Enforcement Discretion**
	No	Enforcement Discretion**	Oversight Focus

* "Can the User Independently Review the Basis?" asks whether the function is intended for the purpose of enabling the user to independently review the basis for the recommendations so that it is not the intent that user relies 436 primarily on any such recommendation (part of criterion (4)).

** "Enforcement Discretion" indicates that, based on our current understanding of the risks of these devices, FDA does not intend at this time to enforce compliance with applicable device requirements.

Table 1 Summary of regulatory policy for CDS software functions

Source: Table 3 from Clinical Decision Support Software Draft Guidance for Industry and U.S. Food & Drug Administration Staff, September 2019

The draft FDA guidance provides a large number of examples of the application of this table.

Finally, there are a number of software applications that run on people’s phones or tablets. The FDA published mobile medical apps (MMA) guidance in 2013, which was updated in 2015 and again in 2019.⁷ Like other products, whether or not a mobile application is a regulated device depends on the product’s intended use. If the software is intended to perform a medical device function, then it is a medical device, regardless of the platform.

In addition to the guidance document, the FDA has created several web pages^{15, 16} that discuss MMA and include various examples.


Regulatory requirements: EU

The EU uses the term 'medical device software' (MDSW) instead of SaMD. It defines MDSW as software that is intended to be used, alone or in combination, for a purpose as specified in the definition of 'medical device' in the MDR or IVDR.⁹

The EU uses a different term because:

1. it does not regulate SaMD with functionality that is limited to storage, communication, lossless compression, or simple searching or that is intended for the benefit of populations rather than individuals and
2. contrary to SaMD, software that fulfils a medical purpose but that is also intended to drive or influence the use of a medical device is still considered to be MDSW, whereas, according to the IMDRF notes, SaMD cannot drive a medical device. Qualification as MDSW is regardless of:
 - its location – for example operating in the cloud, on a computer, on a mobile phone or as an additional functionality on a hardware medical device
 - whether the software, in addition, also drives or influences the use of a (hardware) medical device.

If the software is solely intended to drive or influence the use of a hardware medical device, without by itself creating information for a medical purpose, then it is not considered MDSW but nevertheless is covered by the regulation as an accessory for a medical device or IVD medical device or as an integral part or component of a medical device or IVD medical device.



MDSW operates on computing platforms using operating systems and other platforms to access databases, workflow engines, dynamic link libraries, rules engines, etc. Such platforms usually do not qualify as MDSW, except if they contain functionality that qualifies it as a medical device. Consider an image management platform intended to store and communicate images within a hospital network. Such a platform is not considered MDSW if it only stores and communicates information. If, however, the platform contains an application programming interface (API) to provide third-party modules with access to platform tools – for example to segment tubular and sigmoid structures so they can extract vascular trees and detect tumors – then the platform qualifies as a medical device if its toolkit comes with tool-type claims relating to specific medical purposes.

It is noteworthy that the IMDRF, FDA and EU use different definitions for 'medical device'. Despite these differences, the practical interpretation largely overlaps if one ignores the functional and CDS exemptions applied in the two regions. Consider, for example, fertility and contraception apps and apps to support women during pregnancy. Such apps are in scope of the EU definition,⁹ as the purpose refers to 'support and control of conception'. The IMDRF definition¹ refers to 'control of conception' only, but it acknowledges through its notes that some jurisdictions also consider medical devices to be products used for in vitro fertilization or assisted reproduction technologies. The FDA does not refer to any such purpose in its definition of 'medical device',¹⁷ but, through its jurisprudence, it has indicated considering contraception apps to fall under its scope. In 2018, the FDA granted clearance to a mobile app that calculates when a woman is likely to be fertile, and therefore acts as a contraceptive.¹⁸

A case apart are accessibility apps to help people with visual, auditory or speech impairments to engage with the outside world. Such apps are subject to EU MDR, but they are not considered a medical device in the US, nor are such apps considered SaMD, as they do not meet the definition of the IMDRF. Nevertheless, the IMDRF acknowledges through its notes that some jurisdictions regulate *aids for people with disabilities*.¹

In vitro diagnostics

IVD medical devices are a special kind of medical device. They examine specimens derived from the human body – for example to predict treatment response or reactions. Only software in combination with hardware can examine specimens. Most countries consider SaMD that processes information from an IVD to be a normal medical device, rather than an IVD. This is different in the EU, which for historical reasons considers that some SaMD can be considered an IVD medical device. The IMDRF supports this position.

In 2007, the European Commission and the Member States published a consensus paper¹⁹ on IVD kits measuring parameters, which can be used for evaluating the risk of trisomy 21. These kits measure blood markers. They rely on software-only products that are accessories to these kits. Only when used *in combination* with the software could a claim for the detection of trisomy 21 be assigned to the kit.

Over time, this focus on the software as an accessory of the kit was lost and the scope expanded to include certain independent software such as IVD medical devices. This was eventually 'carved into stone' with the new definition for 'IVD medical device', which refers to software used *alone* for the examination of specimens.²⁰ MDCG guidance tells us that we need to consider the intended purpose and the source of the input data to make the medical device or IVD medical device determination. As many of the purposes listed for medical devices and IVD medical devices overlap – for example 'diagnosis' (medical device) versus 'information concerning a physiological or pathological process or state' (IVD medical device) – and as the source of the input data can be variable (see Figure 3), this creates a rather nebulous legal status for SaMD using IVD input data in the EU.



Figure 3 Example of different medical device/IVD medical device input data used for providing diagnostic information. Traditionally, diabetic patients measured their blood glucose level through a finger prick using an IVD hardware device to analyse the sample. Nowadays, many use medical devices rather than IVD medical devices because they are less cumbersome. Either they wear a monitoring patch on their upper arm or they squeeze a small radio-wave instrument between their fingers. A manufacturer of SaMD that uses blood glucose as input data can no longer be sure whether the data originates from a medical device or an IVD medical device. As the source of the input data is a deciding criterion, this creates legal uncertainty regarding whether the software is a medical device or an IVD medical device. (Cobbaert, 2020)


This situation is not future-proof. Increasingly, clinical decision support systems are being deployed using all kinds of information found in the electronic patient record where it is not clear whether it was derived through a medical device or an IVD medical device and, if it is information from an IVD medical device, to what extent it substantially influences the artificial intelligence (AI) decision for a particular patient. The distinction is important because medical devices and IVD medical devices follow a different regulation in the EU and might require a different Notified Body.

Classification

In the US, a manufacturer typically classifies its SaMD by browsing through FDA databases²¹ to determine the applicable product code and matching device class. If no code appears to fit, the manufacturer can submit a request for information to the FDA²² – a straightforward approach that leads to Class I, II or III classification, requiring, respectively, a registration, a 510(k) submission/De Novo submission or a Premarket Assessment (PMA).

Classification in the EU is more complex. The EU MDR distinguishes Class I, IIa, IIb and III, whereas the EU IVDR uses letters to distinguish the classes: Class A, B, C and D. The MDR also makes a distinction for Class I devices that contain a measuring function, are reusable surgical instruments or are sterile. Only Class I devices that are not sterile, not reusable surgical instruments, and do not contain measuring functionality and Class A IVDs do not require a Notified Body. The class also affects the Notified Body sampling frequency (see Figure 4) and the type and frequency of reporting. In addition, all medical devices require a clinical evaluation²³ or a performance evaluation.²⁴

Many devices and almost all Class III devices and Class IIb implantable devices will require a clinical investigation. Not carrying out a clinical investigation requires a documented justification.



*generic device group: a set of devices having the same or similar intended purposes or a commonality of technology allowing them to be classified in a generic manner not reflecting specific characteristics

Figure 4 Illustration of extent of Notified Body review for increasing device classification. In the EU, a Notified Body increases the technical documentation sampling frequency for higher-class devices. (Cobbaert, 2020)

Classification according to EU IVDR is relatively straightforward, as the classification rules act as a waterfall mechanism (see Figure 5) with relatively few terms that require interpretation. Classification according to EU MDR, on the other hand, is more complex, especially when software is involved.




Figure 5 EU IVDR classification system. The EU IVDR classification rules are ordered from high to low class, making them usable as a waterfall mechanism, in that the first rule that applies to your device will provide for the highest class. This is in contrast to the EU MDR, where, for example, the last rule (i.e. Rule 22) leads to the highest class, Class III. (Cobbaert, 2020)

Under the EU MDR, MDSW is classified through classification Rules 11, 15 and 22, unless it is intended to drive or influence a hardware medical device, in which case other classification rules come into play (outside of the scope of this paper). Note that under the EU MDD, Rules 9 and 10 have also been used to classify independent MDSW, but the MDCG guidance,⁹ by not describing these identically worded rules in the EU MDR, hints at regulators not considering them applicable to independent MDSW under the EU MDR.

Rule 15 applies to contraception apps and leads to Class IIb. Rule 22 leads to Class III and applies to active therapeutic devices with an integrated or incorporated diagnostic function, which significantly determines the patient management. To our knowledge, no SaMD currently exists on the market that is subject to Rule 22, although we expect future digital therapeutics with an integrated diagnostic function (see Figure 6).



Figure 6 Digital doc.²⁵ The avatar will see you now. Virtual therapists used the video game 'Second Life' to reach patients. Although the game involved humans behind the avatar, technology could be developed to make the avatars fully autonomous, similar to personal digital assistants like Siri and Alexa. (Wootton, 2010)

Contrary to Rules 15 and 22, all SaMD in scope of EU MDR is subject to classification Rule 11. The following (italicized) text is verbatim text from the EU MDR. For the purposes of identification in this white paper, the paragraphs have been assigned 'a', 'b', and 'c'. These do not appear in the EU MDR.

[Rule 11a] Software intended to provide information which is used to take decisions with diagnosis or therapeutic purposes is classified as class IIa, except if such decisions have an impact that may cause:

- death or an irreversible deterioration of a person's state of health, in which case it is in class III; or*
- a serious deterioration of a person's state of health or a surgical intervention, in which case it is classified as class IIb.*

[Rule 11b] Software intended to monitor physiological processes is classified as class IIa, except if it is intended for monitoring of vital physiological parameters, where the nature of variations of those parameters is such that it could result in immediate danger to the patient, in which case it is classified as class IIb.

[Rule 11c] All other software is classified as class I.

Rule 11a refers to software intended to provide information that is used to take decisions. These decisions can be made by the software itself, by other software, by hardware or by humans. Rule 11a is applicable not only to software that provides diagnosis or therapy by itself but to all MDSW if this software provides information for a specific medical purpose, as listed in the definition of 'medical device':

- diagnosis, prevention, monitoring, prediction, prognosis, treatment or alleviation of disease
- diagnosis, monitoring, treatment, alleviation of, or compensation for, an injury or disability
- investigation, replacement or modification of the anatomy or of a physiological or pathological process or state ...

Rule 11a does not apply to firmware needed to provide the output of a thermistor in a digital thermometer. That firmware is neither SaMD nor MDSW, as it does not in itself create information for medical purposes (unless the software creates, for example, an alert when fever occurs; such an alert is considered information for medical purposes and such software is considered SaMD and MDSW). Rule 11a also does not apply to fertility apps. Fertility apps are intended to *support conception*, a purpose that qualifies for a medical device, but that is not listed in the definition of 'medical device' as a specific medical purpose. Although they are not described in the MDCG guidance,⁹ it could be argued that accessibility apps intended to *alleviate or compensate for a disability* are not subject to Rule 11a because such apps provide information for very different purposes than specifically medical (see Figure 7). "Accessibility apps are subject to Rule 11c, leading to Class I, unless other rules apply".




Figure 7 Illustration of an application intended to alleviate or compensate for a disability. The app uses brainwaves to allow patients with locked-in syndrome to communicate with the outside world. Locked-in syndrome is a state of wakefulness and awareness with quadriplegia and paralysis of the lower cranial nerves, resulting in inability to show facial expression, move, speak or communicate, except by coded eye movements or through these apps. (Cobbaert, 2020)

If Rule 11a is interpreted literally, all SaMD is Class III because it is always possible, with some imagination, to think of a situation where the information provided by the software leads to death, no matter how unlikely. Instead, regulators advise interpreting Rule 11a using the IMDRF SaMD risk framework, which relies on the significance of the information and the criticality of the disease, condition or intervention (see Figure 8 and Figure 9).










		Significance of Information			
		<ul style="list-style-type: none"> Treat Provide therapy to a human body Diagnose disease/condition Detect disease/condition Screen disease/condition 	<ul style="list-style-type: none"> Aid in treatment Provide enhanced support for safe and effective use of medicinal products or medical device Aid to make a definitive diagnosis Triage or identify early signs of a disease or condition 	<ul style="list-style-type: none"> Inform of options for <ul style="list-style-type: none"> Treatment Diagnosis Prevention Aggregate relevant clinical information 	
Disease Type Patient Condition	Intervention Type	Treat or Diagnose	Drive Clinical Management	Inform Clinical Management	
<ul style="list-style-type: none"> Life-treating 	<ul style="list-style-type: none"> Requires major therapeutic interventions Sometimes time critical Accurate and/or timely diagnosis vital to: avoid death; serious deterioration of health or to mitigate public health risk 	Critical	 Type IV.i	 Type III.i	 Type II.i
<ul style="list-style-type: none"> Moderate in progression Often curable 	<ul style="list-style-type: none"> Does not require major therapeutic interventions Not expected to be time critical Vital to avoiding unnecessary interventions 	Serious	 Type III.ii	 Type II.ii	 Type I.ii
<ul style="list-style-type: none"> Slow with predictable progression of disease state Minor chronic illnesses or states May not be curable 	<ul style="list-style-type: none"> Can be managed effectively 	Non-Serious	 Type II.iii	 Type I.iii	 Type I.i

Figure 8 Illustration of how EU MDR Rule 11a maps on the IMDRF SaMD risk framework. (Cobbaert, 2020)

The second paragraph (Rule 11b) refers to software for monitoring of physiological processes (not just vital physiological processes!).

'Monitoring' refers to monitoring over time or verifying whether a signal is within range – that is checking against limits. As monitoring is a specific medical purpose, MDSW for monitoring of physiological processes is subject to both Rule 11a and Rule 11b.

		Significance of Information		
		<ul style="list-style-type: none"> Treat Provide therapy to a human body Diagnose disease/condition Detect disease/condition Screen disease/condition 	<ul style="list-style-type: none"> Aid in treatment Provide enhanced support for safe and effective use of medicinal products or medical device Aid to make a definitive diagnosis Triage or identify early signs of a disease or condition 	<ul style="list-style-type: none"> Inform of options for <ul style="list-style-type: none"> Treatment Diagnosis Prevention Aggregate relevant clinical information
Disease Type Patient Condition	Intervention Type	Treat or Diagnose	Drive Clinical Management	Inform Clinical Management
<ul style="list-style-type: none"> Life-treating 	<ul style="list-style-type: none"> Requires major therapeutic interventions Sometimes time critical Accurate and/or timely diagnosis vital to: avoid death; serious deterioration of health or to mitigate public health risk 	Critical	<ul style="list-style-type: none"> image-based stroke detection melanoma detection paediatric meningitis detection screening of mutable pathogen 	<ul style="list-style-type: none"> virtual colonoscopy melanoma tracking stroke prediction identifying genetic predisposition to develop sepsis in general population
<ul style="list-style-type: none"> Moderate in progression Often curable 	<ul style="list-style-type: none"> Does not require major therapeutic interventions Not expected to be time critical Vital to avoiding unnecessary interventions 	Serious	<ul style="list-style-type: none"> sleep apnoea detection tinnitus sound therapy diagnoses Parkinson's disease based on data captured by vibration/position sensors 	<ul style="list-style-type: none"> heart arrhythmia detection cardiovascular surgical planning guide for diagnosis of kidney function disorders and cardiac risk insulin dose calculator data collection to guide exercise-based treatment or cardiac rehabilitation patients prediction of asthma episode
<ul style="list-style-type: none"> Slow with predictable progression of disease state Minor chronic illnesses or states May not be curable 	<ul style="list-style-type: none"> Can be managed effectively 	Non-Serious	<ul style="list-style-type: none"> skin lesion tracking to diagnose or rule out eczema 	<ul style="list-style-type: none"> Nystagmus and other eye movement disorder detection provides options for diagnosing seasonal allergic rhinitis vs. common cold alert doctor of potential triggers indicative of cholesterol management issues advise on eff seasonal allergy drug

Figure 9 Examples of SaMD mapped against IMDRF risk framework. These (abbreviated) examples are provided by IMDRF and the FDA¹¹. (Cobbaert, 2020)

As Rule 11a is applicable to nearly all MDSW and leads to Class IIa, IIb or III, but not Class I, the implication is that the vast majority of MDSW is Class IIa or higher and requires a conformity assessment by a Notified Body. In addition, under the EU IVDR, the vast majority of MDSW now requires conformity assessment carried out by a Notified Body. This is in stark contrast to the situation under the existing EU Medical Devices Directives, where only a small fraction of software requires a Notified Body.

Possible future directions

In the US, the FDA is examining alternative regulatory pathways for software. For example, the FDA has proposed a software 'Precertification (Pre-Cert) Program', which is intended to be a regulatory model that is more streamlined and efficient, resulting in getting product to market and to patients faster than existing methods.

This Pre-Cert Program²⁶ involves focusing primarily on the product developer, instead of focusing primarily on the product itself. If the developer can demonstrate a culture of quality, excellence and responsiveness, then the FDA believes that a streamlined approval process could be allowed.

The FDA is piloting this program exclusively on SaMD products so that the focus can be on software development excellence. The current model proposal includes:

- Excellence Appraisal – the FDA evaluates the software development team for organizational excellence, which would include excellence principles of product quality, patient safety, clinical responsibility, cybersecurity responsibility and proactive culture
- Review Determination – a model where lower-risk devices have a streamlined or perhaps no premarket review requirements
- Streamlined Review – the FDA identifies what would be included in a streamlined review
- Real-World Performance – leveraging performance data to support the regulatory status.

AI and machine learning (ML) systems have the potential to significantly improve health care. Some of these systems even have the ability to learn and improve themselves over time. Just like how people learn from experience, 'adaptive' ML systems can increase their performance as they collect and analyse real-world data about their past performance.

This brings a great opportunity but also a great challenge – how to manage timely changes to software that impact product performance? How change of AI during runtime relates to its conformity assessment is a domain in which IMDRF should provide further guidance. The FDA has already published a discussion paper that focuses on this very issue, and cites the Pre-Cert Program as a possible regulatory pathway.


Conclusion

Marc Andreessen, Co-Founder of Netscape, an American company famous for its now obsolete web browser, once said that software is eating the world. He meant that an awful lot of successful technology companies find themselves in a different market than the one they started out in. Today they realize that in order to survive they need to turn themselves into a software company.

Placing SaMD on the market has become significantly harder in the EU, whereas the US has removed some of its 'red tape'. In the EU, SaMD classification has become complex, whereas in the US it is very straightforward.

The health care field is moving faster than it has in the past, and new applications might make us pause and go back to our fundamental goals of assuring safety and effectiveness and discovering alternative paths to reach those goals. The shift from a purely product focus to a product+process viewpoint is a new pathway for the FDA, through which it converges towards the quality management system approach used within the EU.

Product developers are innovators. With the explosion of wearables and objects that are part of the Internet of Things (IoT), health and wellness information and technology can be found everywhere. Such apps could run on the computing platform of your mobile phone but also on your fridge or your car. Technology surrounds us to the point that we suggest that humanity has entered an era of 'everywhereables', and this technology will vastly improve our understanding of the human body. Most of these innovations will be driven by software, and most of that software will be SaMD.



References

- ¹ International Medical Device Regulators Forum (2013) 'Software as a Medical Device (SaMD): Key Definitions'. Available at: <http://www.imdrf.org/docs/imdrf/final/technical/imdrf-tech-131209-samd-key-definitions-140901.pdf> (Accessed 1 September 2020)
- ² Philipson, L. Medical Products Agency Sweden (2008) 'Information Systems and Medical Devices. A classification and borderline issue'
- ³ U.S. Food & Drug Administration (2019) 'Clinical Decision Support Software: Draft Guidance for Industry and Food and Drug Administration Staff'. Available at: <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/clinical-decision-support-software> (Accessed 1 September 2020)
- ⁴ Regulation (EU) 2017/745 on Medical Devices. Available at: <https://eur-lex.europa.eu/legal-content/EN/TXT/?uri=CELEX:02017R0745-20170505> (Accessed 1 September 2020)
- ⁵ Regulation (EU) 2017/746 on In Vitro Diagnostic Devices. Available at: <https://eur-lex.europa.eu/legal-content/EN/TXT/?uri=CELEX:32017R0746> (Accessed 1 September 2020)
- ⁶ In EU MDR Art. 2(11), 'system' means a combination of products, either packaged together or not, which are intended to be inter-connected or combined to achieve a specific medical purpose
- ⁷ U.S. Food & Drug Administration (2019) 'Policy for Device Software Functions and Mobile Medical Applications: Guidance for Industry and Food and Drug Administration Staff'. Available at: <https://www.fda.gov/media/80958/download> (Accessed 1 September 2020)
- ⁸ International Medical Device Regulators Forum (2014) "'Software as a Medical Device": Possible Framework for Risk Categorization and Corresponding Considerations'. Available at: <http://www.imdrf.org/docs/imdrf/final/technical/imdrf-tech-140918-samd-framework-risk-categorization-141013.pdf> (Accessed 1 September 2020)
- ⁹ European Commission (2019) 'Guidance on Qualification and Classification of Software in Regulation (EU) 2017/745 – MDR and Regulation (EU) 2017/746 – IVDR'. Available at: <https://ec.europa.eu/docsroom/documents/37581> (Accessed 1 September 2020)
- ¹⁰ The European Court issued a decision in the case C329/16 Snitem and Philips France vs the French State that the functionality involved in the drug–drug interaction checks qualifies as a medical device
- ¹¹ U.S. Food & Drug Administration (2019) 'Guidance Clinical Decision Support Software: Draft Guidance for Industry and Food and Drug Administration Staff'. Available at: <https://www.fda.gov/media/109618/download> (Accessed 1 September 2020)
- ¹² Unique Device Identification System: A final Rule by FDA. Published 9/24/2013. Docket #: FDA-2011-N-0090. Document Citation: 78FR 58785 pages 58785-58828
- ¹³ U.S. Food & Drug Administration (2019) 'Medical Device Data Systems'. Available at: <https://www.fda.gov/medical-devices/general-hospital-devices-and-supplies/medical-device-data-systems> (Accessed 1 September 2020)
- ¹⁴ HealthIT.gov (2018) 'Clinical Decision Support'. Available at: <https://www.healthit.gov/topic/safety/clinical-decision-support> (Accessed 1 September 2020)

¹⁵ U.S. Food & Drug Administration (2019) 'Examples of Device Software Functions the FDA Regulates'. Available at: <https://www.fda.gov/medical-devices/device-software-functions-including-mobile-medical-applications/examples-device-software-functions-fda-regulates>. (Accessed 1 September 2020)

¹⁶ U.S. Food & Drug Administration (2019) 'Examples of Premarket Submissions that Include MMAs Cleared or Approved by the FDA'. Available at: <https://www.fda.gov/medical-devices/device-software-functions-including-mobile-medical-applications/examples-premarket-submissions-include-mmas-cleared-or-approved-fda> (Accessed 1 September 2020) and (2018) 'FDA allows marketing of first direct-to-consumer app for contraceptive use to prevent pregnancy'. Available at: <https://www.fda.gov/news-events/press-announcements/fda-allows-marketing-first-direct-consumer-app-contraceptive-use-prevent-pregnancy> (Accessed 1 September 2020)

¹⁷ Section 210(h) of the Food, Drug, and Cosmetic Act, United States Code of Federal Regulations, Title 21 Regulation (EU) 2017/746 on In Vitro Diagnostic Devices. Available at: <https://eur-lex.europa.eu/legal-content/EN/TXT/?uri=CELEX:32017R0746> (Accessed 1 September 2020)

¹⁸ U.S. Food & Drug Administration (2018) 'FDA allows marketing of first direct-to-consumer app for contraceptive use to prevent pregnancy'. Available at: <https://www.fda.gov/news-events/press-announcements/fda-allows-marketing-first-direct-consumer-app-contraceptive-use-prevent-pregnancy> (Accessed 1 September 2020)

¹⁹ European Commission (2007) 'Requirements for in vitro diagnostic kits measuring parameters which can be used for evaluating the risk of trisomy 21. Available at: <https://ec.europa.eu/docsroom/documents/10276/attachments/1/translations/en/renditions/native> (Accessed 1 September 2020)

²⁰ Regulation (EU) 2017/746 on In Vitro Diagnostic Devices. Article 2(2). Available at: <https://eur-lex.europa.eu/legal-content/EN/TXT/?uri=CELEX:32017R0746> (Accessed 1 September 2020)

²¹ Search FDA databases by device: <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPCD/classification.cfm> or by 510(k): <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnm.cfm>

²² U.S. Food & Drug Administration (2019) 'FDA and Industry Procedures for Section 513(g) Requests for Information under the Federal Food, Drug, and Cosmetic Act: Guidance for Industry and Food and Drug Administration Staff'. Available at: <http://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM209851.pdf> (Accessed 1 September 2020)

²³ EU MDR Art. 61 and Annex XIV Part A

²⁴ EU IVDR Art. 56 and Annex XIII

²⁵ Schiffman, L. (2010) 'Virtual Therapists: The Avatar Will See You Now'. Popular Science. Available at: <https://www.popsci.com/technology/article/2010-12/avatar-will-see-you-now/> (Accessed 1 September 2020)

²⁶ More information about the Pre-Cert program can be found at <https://www.fda.gov/medical-devices/digital-health/digital-health-software-precertification-pre-cert-program>

Author

Pat Baird works at Philips as Head of Global Software Standards. Pat likes to think of his job as 'policy engineering' – understanding the unmet needs (and frustrations) of regulators and developers, and developing standards, white papers and training to meet those needs. He co-chairs the AdvaMed software committee and an Association for the Advancement of Medical Instrumentation (AAMI) committee on agile software development. His current passion is related to artificial intelligence (AI) in health care; he is a co-chair of multiple AI committees, including for the World Health Organization (WHO), and is an industry representative on a newly formed International Medical Device Regulators Forum (IMDRF) AI committee.

Koen Cobbaert is Senior Manager Quality, Standards and Regulations at Philips. He represents the health industry at the European Commission and IMDRF through, respectively, COCIR and DITTA. He views tech regulations through the computational lens. He enjoys giving training and lectures and thrives on solving complex regulatory challenges. He is interested in AI, wearables and the nature of reality.

Technical reviewers

Jane Edwards is Head of Communications, Global Product Management, at BSI. She holds a BSc in Chemistry and an MBA from Durham University. She has over 13 years' experience in the medical device industry, having previously worked for Coloplast in its ostomy and continence business. Jane's experience includes working within the pharmaceutical, chemical and telecoms industries for Glaxo Wellcome, ICI and Ericsson, allowing her to bring depth of knowledge from many industries and technologies. Her current role at BSI allows her to work with technical reviewers across all disciplines, ensuring that BSI communications are accurate and relevant. She is a member of the European Medical Writers Association.

Eamonn Hoxey is a technical author, trainer and consultant in a range of life science areas, including regulatory compliance, quality management, sterility assurance and standards development. Eamonn worked for Johnson & Johnson (J&J) for 17 years in positions of increasing responsibility for quality and regulatory compliance for medical devices, pharmaceuticals and consumer products. These included Vice President of Compliance, Vice President of Market Quality and Vice President of Quality & Compliance Strategic Programs, leading quality implementation for the EU MDR for J&J's medical devices companies. Prior to joining J&J, Eamonn spent 16 years with the UK Medical Devices Agency, including 6 years as Head of Device Technology and Safety. Eamonn is Chair of CEN/TC 204, Sterilization of medical devices, and chaired ISO/TC 198, Sterilization of health care products, from 2011 to 2019. Eamonn is also a past chair of ISO/TC 210, Quality management and related general aspects for medical devices, and the current Chair of the Board of Directors of AAMI. He received the BSI Wolfe Barry medal in 2016 for his contribution to standards development.

Patricia Kranz-Zuppan is Standards Manager at Medtronic. She has worked in various aspects of the medical device software industry for over 20 years. She is a Medtronic Technical Fellow and Distinguished Engineer responsible for leading and actively participating in the development of international health software standards. Patty is Convener of ISO/TC 210-IEC/SC 62A medical device software joint working group (most notable published standards are IEC 62304 first edition and Amendment 1, IEC TR 80002-1 and IEC TR 80002-2), the co-project leader of IEC 62304 second edition and IEC 82304-1 health software within ISO/TC 215-IEC/SC 62A joint working group 7 for networked health software. Patty is also an active member of ISO/TC 210/JWG 1 on medical device risk management.

Justin McCarthy is Clinical Scientist and Consultant Clinical Engineer for Clin Eng Consulting Ltd. He retired from the NHS as Head of Clinical Engineering in Cardiff in 2009 but continues to provide consultancy services to trusts, the Welsh Government and small businesses. He has been involved in standards at both UK and international levels and served for 7 years until 2018 as Chair of IEC sub-committee SC 62A, responsible for the IEC 60601 series of standards, the primary safety standards for medical electrical equipment. He is one of the co-authors of *Healthcare Technology Management: A Systematic Approach* (Hegarty et al., 2017, CRC Press). He is a Chartered Engineer and a Fellow of the Institution of Engineering and Technology (IET) and of the Institute of Physics and Engineering in Medicine (IPEM).




Paul Sim has worked in the health care industry for over 35 years. He joined BSI in 2010 to lead the organization in Saudi Arabia, where it had been designated as a Conformity Assessment Body. Later, he managed BSI's Unannounced Audit programme. Since October 2015, he has been working with the Notified Body and standards organizations looking at how best to use knowledge, competencies and expertise in both. Previously, he held senior RA/QA leadership positions at Spacelabs Healthcare, Teleflex Medical, Smiths Medical and Ohmeda (formerly BOC Healthcare). Paul is a member of the Association of British Healthcare Industries (ABHI) Technical Policy Group and Convener of the ABHI ISO/TC 210 Mirror Group. He is Convener of the BSI committee that monitors all of the work undertaken by ISO/TC 210, and Convener of the BSI subcommittee dealing with quality systems. As UK Delegation Leader to ISO/TC 210, he is also actively involved in the work of national, European and international standards committees.

Published white papers

- *The proposed EU regulations for medical and in vitro diagnostic devices: An overview of the likely outcomes and consequences for the market*, Gert Bos and Erik Vollebregt
- *Generating clinical evaluation reports: A guide to effectively analysing medical device safety and performance*, Hassan Achakri, Peter Fennema and Itoro Udofia
- *Effective post-market surveillance: Understanding and conducting vigilance and post-market clinical follow-up*, Ibim Tariah and Rebecca Pine
- *What you need to know about the FDA's UDI system final rule*, Jay Crowley and Amy Fowler
- *Engaging stakeholders in the home medical device market: Delivering personalized and integrated care*, Kristin Bayer, Laura Mitchell, Sharmila Gardner and Rebecca Pine
- *Negotiating the innovation and regulatory conundrum*, Mike Schmidt and Jon Sherman
- *The growing role of human factors and usability engineering for medical devices: What's required in the new regulatory landscape?* Bob North
- *ISO 13485: The proposed changes and what they mean for you*, Bill Enos and Mark Swanson
- *The differences and similarities between ISO 9001:2015 and ISO 13485:2016: Can we integrate these quality management standards?*, Mark Swanson
- *How to prepare for and implement the upcoming MDR – Dos and don'ts*, Gert Bos and Erik Vollebregt
- *How to prepare for and implement the upcoming IVDR – Dos and don'ts*, Gert Bos and Erik Vollebregt
- *Planning for implementation of the European Union Medical Devices Regulations – Are You Prepared?*, Eamonn Hoxey
- *Cybersecurity of medical devices: Addressing patient safety and the security of patient health information*, Richard Piggin
- *The European Medical Devices Regulations: What are the requirements for vigilance reporting and post-market surveillance?*, Eamonn Hoxey
- *General Safety and Performance Requirements (Annex 1) in the New Medical Device Regulation: Comparison with the Essential Requirements of the Medical Device Directive and Active Implantable Device Directive*, Laurel Macomber and Alexandra Schroeder
- *Do you know the requirements and your responsibilities for medical device vigilance reporting?: A detailed review on the requirements of MDSAP participating countries in comparison with the European Medical Device Regulation 2017/745*, Cait Gatt and Suzanne Halliday
- *Technical Documentation and Medical Device Regulation: A Guide for Manufacturers to Ensure Technical Documentation Complies with EU Medical Device Regulation 2017/745*, Dr Julianne Bobela, Dr Benjamin Frisch, Kim RoCHAT and Michael Maier

- *Nanotechnology: What does the future look like for the medical devices industry?*, Professor Peter J Dobson, with Dr Matthew O'Donnell
- *Developing and maintaining a quality management system for IVDs*, Melissa Finocchio
- *Digital maturity in an age of digital excitement: Digital maturity goes beyond excitement to quality*, Professor Harold Thimbleby
- *Recent advancements in AI – implications for medical device technology and certification*, Anil Anthony Bharath
- *Risk management for medical devices and the new ISO 14971*, Jos van Vroonhoven
- *The impact and potential for 3D printing and bioprinting in the medical devices industry*, Kenny Dalgarno
- *Sterilization – Regulatory requirements and supporting standards*, Eamonn Hoxey
- *Medical device clinical investigations – What's new under the MDR?*, Maria Donawa

Forthcoming white papers

-  *The convergence of pharma and medical devices*, Barbara Nasto and Jonathan Sutch (working title)
-  *Phthalates and endocrine disruptors*, Benjamin Seery (working title)
-  *Person responsible for regulatory compliance role*, Anne Jury and Maddalena Pinsi

About BSI Group

BSI (British Standards Institution) is the business standards company that equips businesses with the necessary solutions to turn standards of best practice into habits of excellence. Formed in 1901, BSI was the world's first National Standards Body and a founding member of the International Organization for Standardization (ISO). Over a century later it continues to facilitate business improvement across the globe by helping its clients drive performance, manage risk and grow sustainably through the adoption of international management systems standards, many of which BSI originated. Renowned for its marks of excellence including the consumer recognized BSI Kitemark™, BSI's influence spans multiple sectors including aerospace, construction, energy, engineering, finance, healthcare, IT and retail. With over 70,000 clients in 150 countries, BSI is an organization whose standards inspire excellence across the globe. BSI is keen to hear your views on this paper, or for further information please contact us here: julia.helmsley@bsigroup.com

This paper was published by BSI Standards Ltd

For more information please visit:

<http://www.bsigroup.com/en-GB/our-services/medical-device-services/BSI-Medical-Devices-Whitepapers/>



BSI National Standards Body

389, Chiswick High Road
London W4 4AL
United Kingdom

T: +44 (0) 845 086 9001
E: cservices@bsigroup.com
bsigroup.com