

Device Schedule:

Device Name	Model	Type (Codes as per (EU) 2017/2185)	Intended purpose (as per the IFU)	Risk Classification	Basic UDI-DI
<p>Marketing name: ANKORIS</p> <p>Commercial name: ANKORIS</p> <p>Technical name: PODTAY26P</p> <p>Aphakic, Monofocal, Toric, Hydrophilic Acrylic Intraocular Lens</p>	<p>ANKORIS</p>	<p>MDN 1104</p>	<p>The posterior chamber intraocular lens is intended to be placed into the capsular bag with an anterior capsulorhexis for the replacement of the human lens to achieve the visual correction of aphakia in adult patients in whom the cataractous lens has been removed.</p> <p>The lens is indicated for adult patients, with pre-existing astigmatism, surgically treated for cataract, who desire improved uncorrected far vision, with reduced spectacle dependence</p>	<p>Class IIb, Implantable</p>	<p>05414846010030077</p>

Device Name	Model	Type (Codes as per (EU) 2017/2185)	Intended purpose (as per the IFU)	Risk Classification	Basic UDI-DI
<p>Marketing name: FINEVISION</p> <p>Commercial name: POD F</p> <p>Technical name: PODFY26P</p> <p>Aphakic, Multifocal (Trifocal Diffractive), Aspherical, Hydrophilic Acrylic Intraocular Lens</p>	POD F	MDN 1104	<p>The posterior chamber intraocular lens is intended to be placed into the capsular bag with an anterior capsulorhexis for the replacement of the human lens to achieve the visual correction of aphakia in adult patients in whom the cataractous lens has been removed.</p> <p>The lens is indicated for adult patients, surgically treated for cataract, with possibly associated presbyopia, who desire improved uncorrected far vision, useful near and intermediate visual functions and reduced spectacle dependence.</p>	Class IIb, Implantable	05414846010030077

Device Name	Model	Type (Codes as per (EU) 2017/2185)	Intended purpose (as per the IFU)	Risk Classification	Basic UDI-DI
<p>Marketing name: FINEVISION TORIC</p> <p>Commercial name: POD FT</p> <p>Technical name: PODFTY26P</p> <p>Aphakic, Multifocal (Trifocal Diffractive), Toric, Hydrophilic Acrylic Intraocular Lens</p>	POD FT	MDN 1104	<p>The posterior chamber intraocular lens is intended to be placed into the capsular bag with an anterior capsulorhexis for the replacement of the human lens to achieve the visual correction of aphakia in adult patients in whom the cataractous lens has been removed.</p> <p>The lens is indicated for adult patients, with pre-existing astigmatism, surgically treated for cataract, with possibly associated presbyopia, who desire improved uncorrected far vision, useful near and intermediate visual functions and reduced spectacle dependence.</p>	Class IIb, Implantable	05414846010030077

Additional Information:

ANKORIS: Spherical equivalent power at IOL plane (diopter increments): +6D to +30D (0.5D steps)
Cylinder power IOL plane (diopter increments): 1.50D; 2.25D; 3.00D; 3.75D; 4.50D; 5.25D; 6.00D

POD F: Spherical equivalent power at IOL plane (diopter increments): +6D to +35D (0.5D steps)
Additional Power at IOL plane (diopters): +3.50D and +1.75D at IOL plane

POD FT: Spherical equivalent power at IOL plane (diopter increments): +6D to +35D (0.5D steps)
Cylinder power IOL plane (diopter increments): 1.00D; 1.50D; 2.25D; 3.00D; 3.75D; 4.50D; 5.25D; 6.00D
Additional Power at IOL plane (diopters): +3.50D and +1.75D