

Device Schedule:

Device Name	Model	Type (Codes as per (EU) 2017/2185)	Intended purpose (as per the IFU)	Risk Classification	Basic UDI-DI
<p>Marketing name: FINEVISION</p> <p>Commercial name: MICRO F</p> <p>Technical name: MICFY25P</p> <p>Biconvex aspheric trifocal diffractive intraocular lens</p>	<p>MICRO F</p>	<p>MDN 1104</p>	<p>The posterior chamber intraocular lens is intended to be placed into the capsular bag with an anterior capsulorhexis for the replacement of the human lens to achieve the visual correction of aphakia in adult patients in whom the cataractous lens has been removed.</p> <p>The lens should be used as intended in adult patients, surgically treated for cataract, with possibly associated presbyopia, who desire improved uncorrected far vision, useful near and intermediate visual functions and reduced spectacle dependence.</p>	<p>Class IIb, Implantable</p>	<p>0541484601004007C</p>

Additional Information:

MICRO F: Spherical equivalent power at IOL plane (diopter increments): +10D to +35D (0.5D steps)

Additional Power at IOL plane (diopters): +3.50D and +1.75D