

IVDR Classification rules under the IVDR

All devices will need to be divided into classes, A, B, C, or D, taking into account their purpose intended by the manufacturer and inherent risks. The Classification map below allows you to allocate your device correctly. Application of the classification rules shall be governed by the intended purpose, novelty, complexity and inherent risk of the devices.

