

Device Schedule:

Intended Purpose as per the Instructions for Use:

For Use with the cobas® liat CT/NG/MG nucleic acid test:

The **cobas® liat** CT, NG and MG control kit is intended for assay tube Lot Validation with the **cobas® liat** CT/NG/MG nucleic acid test.

For Use with the cobas® liat CT/NG nucleic acid test:

The **cobas® liat** CT, NG and MG control kit is intended for assay tube Lot Validation with the **cobas® liat** CT/NG nucleic acid test.

Device Name	Model	Type (Codes as per (EU) 2017/2185)	Risk Classification	Basic UDI-DI
cobas® liat CT, NG and MG control kit	09449639190	IVR 0503	Class C near-patient test	7613336030039Y