

Device Schedule:

Intended Purpose as per the Instructions for Use:

For use with **cobas® liat** SARS CoV2, Influenza A/B & RSV nucleic acid test:

cobas® liat SARS-CoV-2, Influenza A/B & RSV control kit is intended for assay tube Lot Validation with the **cobas® liat** SARS-CoV-2, Influenza A/B & RSV nucleic acid test.

For use with **cobas® liat** SARS-CoV-2 & Influenza A/B v2 nucleic acid test:

cobas® liat SARS-CoV-2, Influenza A/B & RSV control kit is intended for assay tube Lot Validation with the **cobas® liat** SARS-CoV-2 & Influenza A/B v2 nucleic acid test.

For use with **cobas® liat** SARS-CoV-2 v2 nucleic acid test:

cobas® liat SARS-CoV-2, Influenza A/B & RSV control kit is intended for assay tube Lot Validation with the **cobas® liat** SARS-CoV-2 v2 nucleic acid test.

Device Name	Model	Type (Codes as per (EU) 2017/2185)	Risk Classification	Basic UDI-DI
cobas® liat SARS-CoV-2, Influenza A/B & RSV control kit	09731270190	IVR 0503	Class B near-patient test	761333603138AQ